



Product Service

CERTIFICATE

No. Z2 17 03 13890 02823

Holder of Certificate: **Astec International Ltd.**
16th Floor, Lu Plaza, 2 Wing Yip Street
Kwun Tong
Kowloon
HONG KONG



Certification Mark:



Product: **Power supply
(Switching Power Supply)**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 682301405809

Valid until: 2019-06-19

Date, 2017-03-31

(Yager Bi)

Page 1 of 2



**CERTIFICATE****No. Z2 17 03 13890 02823****Model(s):**

LCC600-28U-4P, LCC600-28U-9P, LCC600-28U-4XXX,
 LCC600-28U-9XXX, LCC600-48U-4P, LCC600-48U-9P,
 LCC600-48U-4XXX, LCC600-48U-9XXX, LCC600-36U-4P,
 LCC600-36U-9P, LCC600-36U-4XXX, LCC600-36U-9XXX,
 LCC600-12U-4P, LCC600-12U-9P, LCC600-12U-4XXX,
 LCC600-12U-9XXX

(where XXX can be any alphanumeric character, symbol or blank that represents customer identity that do not affect safety)

Parameters:

Rated Input : 100-240VAC, 8A MAX, 50/60Hz
 Rated DC Output : For LCC600-28U series:
 +28V, 25A MAX., +5Vsb, 1.5A MAX.
 For LCC600-48U series:
 +48V, 12.5A MAX., +5Vsb, 1.5A MAX.
 For LCC600-36U series:
 +36V, 16.7A MAX., +5Vsb, 1.5A MAX.
 For LCC600-12U series:
 +12V, 50A MAX., +5Vsb, 1.5A MAX.
 For all models:
 Max. Total Output Power: 600W

Construction : Built-in
 Protection Class : I
 Degree of Protection : IPX0

Remarks:

- When installing the equipment, all requirements of the mentioned standard must be fulfilled.
- Refer to the installation and operating instruction from manufacturer for the details of loading condition and operating temperature.
- Clearance was evaluated for operating altitude up to 5000m above sea level.
- These power supplies contains output exceeding 240VA, when installing into end system, care must be taken that the output and associated wire(s) may not be touched.
- Built-in type equipment, suitable enclosure should be provided in end system.
- These power supplies have been evaluated according to EN 60601-1:2006/A1:2013 with the following conditions:
 1. The output was not evaluated as patient connected circuits.
 2. Compliance with the requirements for EMC shall be evaluated for the end use product.
 3. These products have been investigated only as a component part for use in equipment where the suitability of the combination is subject to end product investigation.
 4. These power supplies are designed to be protectively earthed. Earthing connection and continuity test shall be checked in end product.
 5. The leakage current test shall be checked in end product.
 6. The risk management requirements of the standard were not addressed.
 7. Clearance/creepage distance and dielectric strength were evaluated and fulfilled the requirements for MOPP.

Tested according to:

EN 60601-1:2006/A1:2013
 EN 60950-1:2006/A2:2013

Production Facility(ies):

28532

Page 2 of 2