



Ref. Certif. No.

SG-MD-00450M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product  
Produit

Switching power supply unit  
(Switching Mode Power Supply for Building-in)

Name and address of the applicant  
Nom et adresse du demandeur

Astec International Limited - Philippine Branch  
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood  
City Cyberpark, Bagumbayan  
1110 Quezon City, PHILIPPINES

Name and address of the manufacturer  
Nom et adresse du fabricant

Astec International Limited - Philippine Branch, 3rd & 4th Floor, Techno  
Plaza One Bldg, #18 Orchard Road, Eastwood City Cyberpark,  
Bagumbayan, 1110 Quezon City, PHILIPPINES

Name and address of the factory  
Nom et adresse de l'usine

Astec Power Philippines Inc., 104 Laguna Boulevard, Laguna Technopark,  
4026 Sta. Rosa Laguna, PHILIPPINES  
For further information please see attachment

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Rated Input : 100-240VAC, 8.5A MAX, 50/60Hz  
Rated DC Output : See test report for details  
Construction : Built-in  
Protection Class : I  
Degree of Protection : IPX0

Model/type Ref.  
Ref. de type

LCM600Q, LCM600W, LCM600Q-T-401,  
LCM600L, LCM600U, LCM600W-T-401,  
LCM600Q-N, LCM600N, LCM600L-N

Additional information (if necessary)  
Information complémentaire (si nécessaire)

Certificate SG-MD-00450 issued on 2014-06-07 is replaced by this version  
due to technical changes. The risk management requirements of the  
standard were not addressed.

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la


IEC 60601-1/A1:2012

as shown in the Test Report Ref. No.  
which forms part of this certificate  
comme indiqué dans le Rapport d'essais numéro  
de référence qui constitue une partie de ce  
certificat

TÜV SÜD PSB Pte Ltd  
211-400622-100

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2015-02-04  
CBS 15 02 51485 01376

  
( Yager Bi )



TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore

**Additional factory information:**


Name and address of the factory  
*Nom et adresse de l'usine*

Astec Electronics (Luoding) Co. Ltd.  
No. 68 Baocheng Road East, Fucheng Luoding,  
527200 Guangdong,  
People's Republic of China

Zhongshan General Carton Box Factory Co., Ltd.  
No.62, Qi Guan Road West, Shiqi District, 528400,  
Zhongshan City, Guangdong,  
People's Republic of China

Astec Power Philippines Inc.,  
Main Avenue Corner Rd "J",  
Cavite Export Processing Zone, Rosario,  
Cavite, Philippines

Date, 2015 02 04  
CBS 15 02 51485 01376

  
Signature : ( Yager Bi )

TÜV SÜD PSB Pte Ltd • 1 Science Park Drive • Singapore 118221

