



Product Service


CERTIFICATE

No. Z2 15 10 13890 02471

Holder of Certificate: **Astec International Ltd.**
 16th Floor, Lu Plaza, 2 Wing Yip Street
 Kwun Tong
 Kowloon
 HONG KONG



Certification Mark:




Product: **Switch mode power supplies
 (Switching Mode AC-DC Adaptor)**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 686101000404

Valid until: 2020-10-26



Date, 2015-11-02
 Page 1 of 2

(Jimmy Huang)



Product Service

CERTIFICATE
No. Z2 15 10 13890 02471

Model(s): **DPS54-M, DPS55-M, DPS58-M**

Parameters:

Rated DC Input :	100-240Vac, 50/60Hz, 2A
Rated DC Output :	15V, 4A, 60W (for model DPS54-M) 24V, 2.5A, 60W (for model DPS55-M) 48V, 1.25A, 60W (for model DPS58-M)
Protection Class :	I
Degree of Protection :	IP20

Remarks :

- When installing these equipments, all requirements of the mentioned standard must be fulfilled.
- Clearance was evaluated for operating altitude up to 3000m above sea level.
- The maximum operating temperature is 40°C.
- These power supplies have been evaluated according to EN 60601-1:2006/A1:2013 with the following conditions:
 - 1.The output was not evaluated as patient connected circuits.
 - 2.Compliance with the requirements for EMC shall be evaluated for the end use product.
 - 3.These power supplies have been investigated only as a component part for use in equipment where the suitability of the combination is subject to end product investigation.
 - 4.These power supplies are designed to be protectively earthed. Earthing connection and continuity test shall be checked in end product.
 - 5.These power supplies must be installed in accordance with the instruction manual.
 - 6.The leakage current test shall be checked in end product.
 - 7.The risk management requirements of the standard were not addressed.

Tested according to: EN 60950-1:2006/A2:2013
EN 60601-1:2006/A1:2013

Production Facility(ies): 51369

Page 2 of 2