



Ref. Certif. No.

SG-MD-00155A1/M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE)METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switch mode power supplies
(AC/DC Switch Mode Power Supply)

Name and address of the applicant
Nom et adresse du demandeur

Astec International Limited - Philippine Branch
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood
City Cyberpark, Bagumbayan
1110 Quezon City, PHILIPPINES

Name and address of the manufacturer
Nom et adresse du fabricant

Astec International Limited - Philippine Branch, 3rd & 4th Floor, Techno
Plaza One Bldg, #18 Orchard Road, Eastwood City Cyberpark,
Bagumbayan, 1110 Quezon City, PHILIPPINES

Name and address of the factory
Nom et adresse de l'usine

Astec Electronics (Luoding) Co. Limited, No. 68 Baocheng Road East,,
Fucheng Luoding,, 527200 Guangdong, PEOPLE'S REPUBLIC OF CHINA
For further information please see attachment

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Rated Input : 100-240VAC, 0.3A, 50/60Hz
Rated Output : 12VDC, 1.0A max., 12W max.
Protection Class : II
Degree of Protection : IP40 [for Model DA12-120MP-M]
IP42 [for other Models]

Model/type Ref.
Ref. de type

DA12-120EU-M, DA12-120UK-M, DA12-120MP-M,
DA12-120US-M, DA12-120AU-M

Additional information (if necessary)
Information complémentaire (si nécessaire)

Certificate SG-MD-00155A1 issued 2012-02-27 is replaced by this version
due to technical changes.

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la


IEC 60601-1:2005

as shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificat

TÜV SÜD PSB Pte Ltd
211-300650-101

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2015-01-26
CBS 15 01 51485 01373


(Yager Bi)



TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore

Additional factory information:

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc.,
Main Avenue Corner Rd "J",
Cavite Export Processing Zone, Rosario,
Cavite, Philippines

Date, 2015-01-26
CBS 15 01 51485 01373

Signature :  (Yager Bi)