



Ref. Certif. No.

SG-MD-00126M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE)METHODE OC**CB TEST CERTIFICATE**
CERTIFICAT D'ESSAI OCProduct
ProduitSwitch mode power supplies
(Switch Mode Power Supply for Building-in)Name and address of the applicant
Nom et adresse du demandeurAstec International Limited - Philippine Branch
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood
City Cyberpark, Bagumbayan
1110 Quezon City, PHILIPPINESName and address of the manufacturer
Nom et adresse du fabricantAstec International Limited - Philippine Branch, 3rd & 4th Floor, Techno
Plaza One Bldg, #18 Orchard Road, Eastwood City Cyberpark,
Bagumbayan, 1110 Quezon City, PHILIPPINESName and address of the factory
Nom et adresse de l'usineAstec Power Philippines Inc., 104 Laguna Boulevard, Laguna Technopark,
4026 Sta. Rosa Laguna, PHILIPPINES
For further information please see attachmentRatings and principal characteristics
Valeurs nominales et caractéristiques principalesRated AC Input : 100-240V, 1.8A, 50/60Hz
Input power must be less than 74W
Rated DC Output : See test report for details
Protection Class : I
Degree of Protection : IPX0Trade mark (if any)
Marque de fabrique (si elle existe)

EMERSON

Model/type Ref.
Ref. de type


NPT42-M, NPT43-M, NPT44-M

Additional information (if necessary)
Information complémentaire (si nécessaire)

See Test Report for National Differences and Group Differences

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1:2005

as shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificatTÜV SÜD PSB Pte Ltd
211-2311180-100This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Date, 2012-08-17
CBS 12 08 51485 01085
(Jimmy Huang)

TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221



PSB Singapore


Additional factory information:

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc.,
Main Avenue Corner Rd "J",
Cavite Export Processing Zone, Rosario,
Cavite, Philippines

Date, 2012-08-17
CBS 12 08 51485 01085

Signature :


Jimmy Huang